

Choice of balloon and valve size

- Predilatation in all patients, balloon 1-3 mm smaller than perimeter derived annular diameter
- Valve according to perimeter derived annular diameter, oversize in borderline cases
- Postdilatation only if relevant paravalvular regurgitation, with a balloon 1-2 mm smaller than perimeter derived annular diameter

